



Artery Research

ISSN (Online): 1876-4401

ISSN (Print): 1872-9312

Journal Home Page: <https://www.atlantis-press.com/journals/artres>

P1.39: AORTIC PULSE WAVE VELOCITY: SHOULD THE CAROTID – FEMORAL DISTANCE BE MEASURED ON BODY SURFACE OR ESTIMATED FROM BODY HEIGHT?

J. Filipovsky, O. Mayer, M. Dolejsova, J. Seidlerova, L. Bolek, L. Handl

To cite this article: J. Filipovsky, O. Mayer, M. Dolejsova, J. Seidlerova, L. Bolek, L. Handl (2008) P1.39: AORTIC PULSE WAVE VELOCITY: SHOULD THE CAROTID – FEMORAL DISTANCE BE MEASURED ON BODY SURFACE OR ESTIMATED FROM BODY HEIGHT?, Artery Research 2:3, 101–101, DOI: <https://doi.org/10.1016/j.artres.2008.08.346>

To link to this article: <https://doi.org/10.1016/j.artres.2008.08.346>

Published online: 21 December 2019

Germany). The obtained PTT values (74.86 ± 8.63 ms) were compared with PTT evaluated on the same subjects by means of applanation tonometry applied simultaneously on the same locations (75.85 ± 8.61 ms). The two techniques were very well correlated ($r=0.89$, $P<0.001$, Spearman rho 0.88) and values were not statistically different ($p=0.377$). Our preliminary results demonstrate that laser-based non-contact measurement of pulse transit time is feasible in young healthy volunteers, and yields values that are equivalent to those measured using arterial applanation tonometry. Clinical application of this appealing non-invasive method can overcome practical and technical limitations inherent to currently used methods such as arterial applanation tonometry, ultrasound, plethysmography, requiring physical contact of the probe with the patient.

doi:10.1016/j.artres.2008.08.343

P1.37 NON-INVASIVE QUANTITATIVE ASSESSMENT OF ATHEROSCLEROSIS WITH THE PULSE WAVE VELOCITY

I. Orlova, F. Ageev, I. Barinova.
Cardiology Research Center, Moscow, Russian Federation

Arterial stiffness is a predictor of atherosclerosis. This study was conducted to develop a method of coronary atherosclerosis severity assessment by means of brachial-ankle pulse wave velocity (baPWV). MATERIALS AND METHODS. We measured baPWV in 126 males (age 56.2 ± 8.9) who received coronary angiographic examination (CAG). baPWV was measured by Vasera VS-1000 (Fukuda Denshi). RESULTS The patients were categorized into 3 groups according to the number of major coronary arteries having stenosis, that is, 1 vessel disease (1VD) group, 2VD group and 3VD group. The baPWV value was significantly greater in 2VD ($n = 46$, $baPWV = 13.82 \pm 2.40$ m/sec, $p=0.049$) and 3VD groups ($n = 44$, $baPWV = 14.38 \pm 2.97$ m/sec, $p=0.0028$) than that in 1VD group ($n = 36$, $baPWV = 12.49 \pm 2.17$ m/sec). No significant difference was observed between PWV value in 2VD and 3VD groups. To further investigate the relationship between baPWV values and CAG findings, we assessed the severity of stenosis (1 group - less than 75% stenosis, 2gr. - 75 to 99% stenosis, and 3gr. - complete occlusion, respectively). The baPWV value was significantly greater in 2 ($n = 56$, $baPWV = 13.84 \pm 2.25$ m/sec, $p=0.025$) and 3 groups ($n = 45$, $baPWV = 14.16 \pm 3.32$ m/sec, $p=0.007$) than that in 1 group ($n = 25$, $baPWV = 12.23 \pm 1.42$ m/sec). No significant difference was observed between baPWV value in 2 and 3 groups. CONCLUSION. baPWV significantly increases with the number of affected vessels and severity of stenosis which indicates that it is a powerful diagnostic instrument for determining coronary artery atherosclerosis in males.

doi:10.1016/j.artres.2008.08.344

P1.38 CONVERSION BETWEEN DEFINITIONS OF PULSE WAVE VELOCITY

S.J. Vermeersch¹, E. Rietzschel², T.C. Gillebert², S. Laurent³, P. Boutouyrie³, P. Segers¹.

¹ *IBiTech, Ghent University, Ghent, Belgium*

² *Dept. Cardiovascular Diseases, Ghent University Hospital, Ghent, Belgium*

³ *Université Paris-Descartes, Faculté de Médecine Paris Descartes, INSERM UMRS872, APHP - HEGP, Dept. Pharmacology, Paris, France*

Background: Different methodologies for determination of carotid-femoral pulse wave velocity (PWV) exist. Distance (L) can be measured from carotid-femoral measurement sites (L_{direct}) or obtained by subtracting carotid-sternum from sternum-femoral distances ($L_{subtracted}$). Transit times are usually obtained either by detection of the maximal upstroke ($\Delta t_{maximal\ upstroke}$) or the foot ($\Delta t_{intersecting\ tangent}$) of the waveform at the measurement sites. This study investigates conversion factors between PWV methodologies.

Methods: 3043 subjects in which both distance measurements were available were divided into model and validation groups (1502/1541 subjects, respectively). In the model population the main determinants of the $ratio_{distance} = L_{subtracted}/L_{direct}$ were determined and a multivariate model was constructed. Estimated $ratio_{distance,est}$ was used to convert from PWV_{direct} to $PWV_{subtracted,est}$ in the validation population. $PWV_{subtracted,est}$ was compared to measured $PWV_{subtracted}$. Ninety three subjects in which both transit times were available were divided into model and validation groups (46/47 subjects, respectively). In the model population a model for estimation of $\Delta t_{maximal\ upstroke,est}$ from $\Delta t_{intersecting\ tangent}$ was constructed and used to estimate $\Delta t_{maximal\ upstroke,est}$ in the validation population. $\Delta t_{maximal\ upstroke,est}$ was compared to measured $\Delta t_{maximal\ upstroke}$. Data are presented as mean(stdev).

Results: The main determinants of $ratio_{distance}$ were age ($R^2=0.17$) and BMI ($R^2=0.15$) (combined: $R^2=0.27$, all $P<0.001$). $PWV_{subtracted,est}$ correlated well with $PWV_{subtracted}$ ($R=0.97$, $P<0.001$) with mean difference of 0.0007 (0.40) m/s. $\Delta t_{maximal\ upstroke,est}$ correlated well with $\Delta t_{maximal\ upstroke,est}$ ($R=0.82$, $P<0.001$) and mean difference of 0.58 (1.03) m/s.

Conclusions: Differences in absolute PWV values are important to compensate for in order to compare between studies. The models proposed allow for such conversion.

doi:10.1016/j.artres.2008.08.345

P1.39 AORTIC PULSE WAVE VELOCITY: SHOULD THE CAROTID – FEMORAL DISTANCE BE MEASURED ON BODY SURFACE OR ESTIMATED FROM BODY HEIGHT?

J. Filipovsky, O. Mayer, M. Dolejsova, J. Seidlerova, L. Bolek, L. Handl.
Charles University, Pilsen, Czech Republic

Objective: Aortic pulse wave velocity (PWV) can be biased by the measurement of carotid – femoral (c-f) distance on body surface. We wondered whether the estimation of distance according to body height could be used.

Methods: Three cohorts of altogether 598 subjects (mean age 58,9 years) were studied. PWV was measured by Sphygmocor device. The c-f distance was 1. measured by tape, 2. estimated: height was multiplied by 0,27 (= median ratio of measured c-f distance to body height).

Results: Difference in PWV calculated by the two methods (measured minus estimated) increased with PWV: it was -0.2 m/s for PWV 5 m/s and +1.8 m/s for PWV 15 m/s. In multiple regression analysis, this difference depended highly significantly ($p<0.0001$) on PWV, weight (positive associations) and height (negative association); there were weak positive associations ($p<0.05$) with male gender, high LDL level and presence of cardiovascular disease and no associations with age, smoking, hypertension or diabetes.

Conclusions: When PWV is estimated from body height, the highest PWV values show regression to the mean. Besides PWV, anthropometric parameters are major determinants of the differences between the two methods. Estimation of c-f distance from body height would simplify the procedure and bias due to obesity and body disproportion would probably be minimized. For future use of aortic PWV, the best method of the distance assessment should be studied in larger cohorts with known cardiovascular morbidity/mortality endpoints.

doi:10.1016/j.artres.2008.08.346

P1.40 DETERMINATION OF PRESSURE INDEPENDENT ARTERIAL STIFFNESS BY CORRECTING PULSE WAVE VELOCITY FOR PRESSURE-AREA RELATIONSHIP

E. Hermeling¹, K.D. Reesink¹, A.A. Kroon¹, P. Segers², A.P.G. Hoeks¹.

¹ *Cardiovascular research institute, Maastricht University, Maastricht, Netherlands*

² *Institute of Biomedical Technology, Ghent University, Ghent, Belgium*

Background: Intrinsic (pressure independent) arterial stiffness is becoming increasingly important as treatment target in hypertensive patients but is still difficult to assess non-invasively. We measured pulse wave velocity (PWV) and the pressure-area (p-A) relation to determine both apparent and intrinsic stiffness.

Method: Non-invasive PWV (nPWV) was measured by multiple-M-mode ultrasound in a phantom. The incisura of the diameter waveforms was used as time-reference point for calculating nPWV. A catheter was placed in the phantom to measure the pressure waveform simultaneously. Additionally, in hypertensive patients carrying a baroreceptor stimulator, finger-pressure and nPWV at the common carotid artery were measured simultaneously. For both phantom and subject studies, intrinsic PWV (PWV_{int}) was derived employing the Bramwell-Hill equation with the incremental distensibility, $dA/(A*dp)$, based on either a linear or exponential p-A relation.

Results: In the phantom setup, nPWV (12.3 ± 0.8 m/s) increased with increasing pressure ($r=0.67$, $p<0.0001$). Because a linear p-A relation was observed, intrinsic PWV was calculated as $PWV_{int}^2 = A_{int}/A*nPWV^2$ and will be independent of pressure ($r=-0.001$). During baroreceptor stimulation MAP decreased from 138 ± 22 to 109 ± 1 mmHg and nPWV decreased from 10.5 ± 1.5 to 6.6 ± 1.3 m/s ($p=0.03$). In these patients the observed p-A relation was exponential and PWV_{int} was therefore calculated using $PWV_{int}^2 = p_{int}/p*nPWV^2$. PWV_{int} did not decrease upon stimulation ($p=0.23$).

Conclusion: Commonly apparent PWV is measured. Intrinsic stiffness can be calculated using nPWV and assuming a linear or exponential p-A relation. An